

# CERTIFICATE

## Berlin Cert

Prüf- und Zertifizierstelle für Medizinprodukte GmbH

certifies, that the following product of the company

has been tested according to DIN EN 149:2009, Section 8.11 – Penetration of filter material in accordance with EN 13274-7 with the listed results:

**Product:** FFP2 Mask Koumask KKF-1A-I

**Requirements:** Penetration of sodium chloride particles: 5,6 %  
 Maximum of 3 masks as received: 5,4 %  
 Maximum of 3 masks after the simulated wearing treatment: 5,6 %  
 Maximum of 3 masks after the test for mechanical strength followed by temperature conditioning: 0,6 %

Penetration of paraffin oil particles: 4,2 %  
 Maximum of 3 masks as received: 2,4 %  
 Maximum of 3 masks after the simulated wearing treatment: 1,3 %  
 Maximum of 3 masks after the test for mechanical strength followed by temperature conditioning: 4,2 %

The penetration of the tested mask is below the maximum permitted penetration of 6% for FFP2 masks.

The conducted tests are documented in the test report P-20-171-G-PA166-5.

This certificate is not a confirmation that the product complies with the german medical device law or the directive 93/42/EEC

Issued on: 2020-12-09

Lennert Lehmann  
authorized signature